



MagQuant™ Plus DNA V2

Catalog Nos. MQP-50010, MQP-50096, MQP-50384, MQP-51920

Manual Revision v1.0

DNA and Library
Normalization Kit

- Magnetic beads based chemistry
- No centrifugation or filtration

PROTOCOL

Contents

Product Description and Process.....	1
Kit Content, Storage, Preparation of Reagents	2
Protocol - Genomic DNA	3
Protocol - PCR Product (amplicons).....	5
Ordering Information	7

For Research Use Only. Not for use in diagnostic procedures.

Information in this document is subject to change without notice.

MAGBIO GENOMICS, INC. DISCLAIMS ALL WARRANTIES WITH RESPECT TO THIS DOCUMENT, EXPRESSED OR IMPLIED, INCLUDING BUT NOT LIMITED TO THOSE OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. TO THE FULLEST EXTENT ALLOWED BY LAW, IN NO EVENT SHALL MAGBIO GENOMICS BE LIABLE, WHETHER IN CONTRACT, TORT, WARRANTY, OR UNDER ANY STATUTE OR ON ANY OTHER BASIS FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH OR ARISING FROM THIS DOCUMENT, INCLUDING BUT NOT LIMITED TO THE USE THEREOF, WHETHER OR NOT FORESEEABLE AND WHETHER OR NOT MAGBIO GENOMICS, INC. IS ADVISED OF THE POSSIBILITY OF SUCH DAMAGES.

TRADEMARKS

The trademarks mentioned herein are the property of MagBio Genomics, Inc. or their respective owners.

Product Description

The MagQuant™ Plus DNA Normalization V2 is a paramagnetic bead-based kit. It is engineered to release the same output of DNA regardless of initial input DNA concentration without the need for fluorescent measurement or other adjustment to obtain the desired uniform DNA concentration from samples of various sources; therefore, saving time and operation costs. The MagQuant™ Plus DNA Normalization V2 is based on binding of DNA to proprietary beads with limited binding capacity; excess DNA is washed off and normalized amounts of DNA are eluted. The protocol requires no centrifugation step; it can be used in manual procedure and well as automatic workflow.

Benefits:

- Equalizing input genomic DNA concentration for DNA libraries construction to help
- produce consistent and reliable NGS data without tedious initial input DNA quantitation.
- No centrifugation or filtration
- Reduce library construction time, reagents usage and overall costs

Applications:

- PCR
- Cloning
- Genotyping
- Target enrichment
- Library Construction
- Next generation sequencing

Process

The workflow of normalization using MagQuant™ Plus DNA V2 consists of 3 simple steps: Bind, Wash, and Elute; that allows the user to obtain equal amounts of DNA output regardless of DNA input. Thus, similar sized PCR DNA fragments (unpurified or purified), purified/unpurified plasmid DNA, as well as DNA from PCR reactions (amplicons) and plasmid lysates can be normalized for various downstream applications such as library preparation, NGS or any other molecular application.

Safety Information

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, please consult the appropriate material safety data sheets (MSDSs). MSDS can be downloaded from the “Product Resource” tab when viewing the product kit.

Kit Content and Storage

Magquant™ Plus DNA V2 Catalog No.	MQP-50010	MQP-50096	MQP-50384	MQP-51920	STORAGE
Number of Preps	10	96	384	1920	
MAG-C7 Particles	110 µL	1 mL	4.5 mL	20 mL	2-8°C
Binding Buffer	750 µL	5 mL	80 mL	100 mL	15-25°C
MB Elution Buffer	1.5 mL	10 mL	40 mL	200 mL	15-25°C

Preparation of Reagents

Catalog No.	Component	Add 100% Isopropanol	STORAGE
MQP-50010	Binding Buffer	75 µL*	15-25°C

Catalog No.	Component	Add 100% Isopropanol	STORAGE
MQP-50096	Binding Buffer	500 µL*	15-25°C

Catalog No.	Component	Add 100% Isopropanol	STORAGE
MQP-50384	Binding Buffer	2 mL*	15-25°C

Catalog No.	Component	Add 100% Isopropanol	STORAGE
MQP-51920	Binding Buffer	10 mL*	15-25°C

*Ensure bottle/tube lid is closed tightly when preparing and storing reagents.

Stability

All components of MagQuant™ Plus DNA Normalization V2 are guaranteed for at least 14 months from the date of purchase when stored as follows: MAG-C7 Particles should be stored at 2-8°C. All other components should be stored at room temperature (15-25°C). Check buffers for precipitates before use. Re-dissolve any precipitates by warming to 37°C.

Specifications and Recommendations

The binding capacity of the bead varies with the size and source of DNA. The amount of DNA that will bind to the beads depends on the efficiency of the extraction protocol, size of the DNA, quality and quantity of the starting material.

Minimum DNA input for genomic DNA normalization is 800 ng and for amplicon normalization is 300 ng.

MagQuant™ Plus DNA V2 Protocol for Genomic DNA

Equipment and Reagents to Be Supplied by User:

- 100% Ethanol
- Magnetic separation device for 1.5 mL tube format or 96 plate format
For 1.5 mL tube format: MagBio Genomics Cat# MBMS-12
For 96 plate format: MagBio Genomics Cat# MYMAG-96X
- 1.5 mL tubes or 96-well cycling plate

Protocol

IMPORTANT: Bring MAG-C7 Particles to room temperature for at least 30 min before use.

1. Transfer 50 µL of genomic DNA to a 96 well plate. If the DNA amount is less than 50 µL adjust the DNA volume to 50 µL with MB Elution Buffer or Nuclease-Free Water.

2. Add 50 µL Binding Buffer and 10 µL of MAG-C7 Particles (shake or vortex MAG-C7 Particles thoroughly to fully resuspend the magnetic beads before pipetting). Mix the DNA sample with the Binding Buffer and the beads thoroughly by pipetting or vortexing. Incubate at RT for 5 min.

3. Remove and discard the supernatant. Do not disturb MAG-C7 Particles while discarding the supernatant.

4. Keep the sample plate on the magnetic separation device and add 150 µl 80% Ethanol to each well and incubate for 1 minute at room temperature.

5. With the sample plate still on the magnetic separation device, remove and discard the cleared supernatant. Do not disturb MAG-C7 Particles while discarding the supernatant.

6. Repeat Steps 4-5 for a second 80% Ethanol wash step.

7. Leave the sample plate on the Magnetic Separation Device for 5 minutes to air dry the MAG-C7 Particles. Remove any residual liquid with a pipette. Do not disturb MAG-C7 Particles.

Note: It is crucial to completely remove all traces of alcohol but take caution in not over drying the beads as this will reduce the yield.

8. Remove the sample plate from the magnetic separation device. Add 25-50µl MB Elution Buffer to each sample and mix thoroughly to resuspend the beads.

9. Seal the plate and incubate for 5 min at 65°C.

10. Place the plate on the magnetic separation device to magnetize the MAG-C7 Particles. Let the plate sit at room temperature until the MAG-C7 Particles are completely cleared from solution.

11. Transfer the supernatant containing the normalized DNA to a new plate.

12. Store the DNA at -20°C.

MagQuant™ Plus DNA Protocol for PCR Product (amplicons)

Equipment and Reagents to Be Supplied by User:

- 100% Ethanol
- Magnetic separation device for 1.5 mL tube format or 96 plate format
For 1.5 mL tube format: MagBio Genomics Cat# MBMS-12
For 96 plate format: MagBio Genomics Cat# MYMAG-96X
- 1.5 mL tubes or 96-well cycling plate

Protocol

IMPORTANT: Bring MAG-C7 Particles to room temperature for at least 30 min before use.

1. To 50 µL of purified PCR, add 50 µL of Binding Buffer and 10 µL of MAG- C7 Particles (shake or vortex MAG- C7 Particles thoroughly to fully resuspend the magnetic beads before pipetting). Mix the PCR amplicon with Binding Buffer and the beads thoroughly by pipetting or vortexing. Incubate at RT for 10 min. If the amplicon amount is less than 50 µL adjust the volume to 50 µL with MB Elution Buffer or Nuclease-Free Water

2. Place the plate on the magnetic separation device to magnetize MAG-C7 Particles. Let the plate sit at room temperature until the MAG-C7 Particles are completely cleared from solution.

3. Remove and discard the supernatant. Do not disturb MAG-C7 Particles while discarding the supernatant.

4. Keep the sample plate on the magnetic separation device and add 150 µl 80% Ethanol to each well and incubate for 1 minute at room temperature

5. With the sample plate still on the magnetic separation device, remove and discard the cleared supernatant. Do not disturb MAG-C7 Particles while discarding the supernatant.

6. Repeat Steps 4-5 for a second 80% Ethanol wash step.

7. Leave the sample plate on the Magnetic Separation Device for 5 minutes to air dry the MAG-C7 Particles. Remove any residual liquid with a pipette. Do not disturb MAG-C7 Particles.

Note: It is crucial to completely remove all traces of alcohol but take caution in not over drying the beads as this will reduce the yield.

8. Remove the sample plate from the magnetic separation device. Add 25-50µl MB Elution Buffer to each sample and mix thoroughly to resuspend the beads.

9. Seal the plate and incubate for 5 min at 65°C.

10. Place the plate on the magnetic separation device to magnetize the MAG-C7 Particles. Let the plate sit at room temperature until the MAG-C7 Particles are completely cleared from solution

11. Transfer the supernatant containing the normalized amplicons to a new plate.

12. Store amplicons at -20°C.

Ordering and Related Product Information

DNA and Library Normalization

Catalog No.	Product
MQP-50010	MagQuant™ Plus DNA Normalization (10 preps)
MQP-50096	MagQuant™ Plus DNA Normalization (96 preps)
MQP-50384	MagQuant™ Plus DNA Normalization (384 preps)
MQP-51920	MagQuant™ Plus DNA Normalization (1920 preps)

Post PCR and Next Gen library prep clean up system

Catalog No.	Product
AC-60005	HighPrep PCR (5 mL)
AC-60050	HighPrep PCR (50 mL)
AC-60250	HighPrep PCR (250 mL)
AC-60500	HighPrep PCR (500 mL)

BigDye Sanger Sequencing Cleanup

Catalog No.	Product
DT-70005	HighPrep DTR (5 mL)
DT-70050	HighPrep DTR (50 mL)
DT-70250	HighPrep DTR (250 mL)
DT-70500	HighPrep DTR (500 mL)

RNA or cDNA for in vitro applications clean up system

Catalog No.	Product
RC-90005	HighPrep RNA Elite (5 mL)
RC-90050	HighPrep RNA Elite (50 mL)
RC-90500	HighPrep RNA Elite (500 mL)

Magnetic Separation Devices

Catalog No.	Description
MYMAG-96	Handheld Magnetic Separation Device (96 well microplate format)
MBMS-10	MagStip magnetic stand (1.5 mL x 10)
MBMS-31550	15ml and 50ml magnetic stand combo. (3x15ml and 3x50ml)



www.magbiogenomics.com